



AURELIA® UNIQUE

Premium Soft Stretch Vinyl Examination Gloves



Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 455 1-4 Medical gloves for single use. BS EN 1186:2002 Materials and articles in contact with foodstuffs.


AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk
+441733 615215

Supermax Healthcare Ltd,
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN

aureliagloves.co.uk



Sizes & Dimensions		Single wall thickness	Physical Properties			
S	80+10 mm	min. 240 mm	 ① min. 0.05 mm ② min. 0.05 mm ③ min. 0.04 mm			
M	95+10 mm	min. 240 mm		Tensile Strength	Unaged Min. 11.0 MPa	Aged Min. 11.0 MPa
L	110+10 mm	min. 240 mm		Elongation	Min. 350%	Min. 350%
XL	> 110 mm	min. 240 mm		Strength at Break	Min. 3.6N	Min. 3.6N

- Stretch Vinyl
- Powder Free
- 3 Years
- Smooth
- 5.5g +/- 0.2g
- Creamy Yellow

